


Services available at the Force Laboratory

Calibration

MENSURAND	MEASURING INTERVAL	EXPANDED UNCERTAINTY ($k=2$)	CMC (CIPM-MRA)
Force	50 N - 5 kN	2×10^{-5}	
	5 kN - 100 kN	2×10^{-5}	
	100 kN - 1 MN	1×10^{-4}	-----

Metrological Control

EQUIPMENT	TESTS	LAWS
Load Cell	Pattern Approval	Not applicable